



Stilvi Pharmaceuticals SA

18 Poseidonos Avenue
GR – 17674 Kallithea / Athen
Greece



Available from
Pharmacy and commerce

Fast, reliable decision-making at the point-of-care



- BfArM listed: Test ID AT376/20
 - reimbursable according to the regulation on entitlement to testing for direct detection of the corona virus SARS-CoV-2
- evaluated by the Paul-Ehrlich-Institut
- CE certified
- Assay Format: Lateral Flow Test
- Sensitivity: 95,56%
- Specificity: 99,58%
- Users: only for professional use
- Storage temperature: 4 – 30 ° C
- Shelf-live: 12 months
- Sample material: nasal swab or nasopharyngeal sample
- Test result in 15 minutes
- Manufacturer: Prognosis Biotech SA, Farsalon, Greece



Stilvi Pharmaceuticals SA

18 Poseidonos Avenue
GR – 17674 Kallithea / Athen
Griechenland



Kit – Equipment	VE 10	VE 30
Re - sealable bag	1	1
Tubes with lateral flow test strips	10	30
Drip bottle with running buffer	1	2
Extraction tubes in transparent films	10	30
Sterile smear swabs in transparent films	10	30
Positive control	1	1
Negative control	1	1
Tube holder	1	1
Quick instructions	1	1
Instructions for use	1	1

Collection Pharmaceutical Wholesale Trade,
Specialized Medical Trade, Specialized Laboratory Trade,
directly from Stilvi Pharmaceuticals SA

Order number PZN: 17557379 (VE 10); PZN: 17557385 (VE: 30)

Product number VE 10: 20.02.02
VE 30: 20.02.01